



ichroma™ Dengue NS1 Ag

INTENDED USE

ichroma™ Dengue NS1 Ag is a fluorescence Immunoassay (FIA) for the qualitative determination of NS1 Antigen in human whole blood/serum/plasma during dengue virus infection. It is useful as an aid in screening of early Dengue virus infection.

For *in vitro* diagnostic use only.

INTRODUCTION

Nonstructural protein 1 (NS1), one of the dengue viral nonstructural proteins, plays a role in supporting the replication complex and attenuating the host immune response against viral infection.¹⁾ Several lines of evidence show that plasma level of secreted NS1 (sNS1) correlated with viremia levels in dengue virus infected patients. The more generated dengue virus, the higher concentration of sNS1 occurred after onset of illness.^{2,3)} The amount of sNS1 is decreased when plasma viral level is reduced. Thus, it is reasonable to detect the sNS1 in patient blood, which makes it possible to early diagnosis of dengue virus infection.

PRINCIPLE

The test uses a sandwich immunodetection method; granulated detector antigen in cartridge binds to antibody in the sample, forming antibody-antigen-antibody complexes, and migrates onto nitrocellulose matrix to be captured by the other immobilized-antigen on the test strip.

The more antibody in sample forms the more antigen-antibody complex, which leads to stronger intensity of fluorescence signal on detector antigen, which is processed by the instrument for ichroma™ tests to show 'dengue NS1 Ag positive' in sample.

COMPONENTS

ichroma™ Dengue NS1 Ag consists of 'Cartridges', 'Detection Buffer Tubes', 'Diluent vial' and an 'ID chip'.

- The cartridge contains a test strip, the membrane which has streptavidin at the test lines, while chicken IgY (cIgY) antibody at the control line.
- Each cartridge is individually sealed in an aluminum foil pouch containing a desiccant. 25 sealed cartridges are packed in a box which also contains an ID chip.
- The detection buffer contains granules. There are contain anti-Dengue NS1 antibody fluorescence conjugate, anti-chicken IgY fluorescence conjugate, bovine serum albumin (BSA) as a stabilizer and sodium azide as a preservative in sodium borate.
- 25 detection buffer tubes are packed in an aluminum foil pouch.
- The diluent contains a detergent and bovine serum albumin (BSA) as a stabilizer and sodium azide in sodium borate buffer as a preservative. The diluent is dispensed in a vial.

WARNINGS AND PRECAUTIONS

- For *in vitro* diagnostic use only.
- Carefully follow the instructions and procedures described in this 'Instruction for use'.
- Use only fresh samples and avoid direct sunlight.
- Lot numbers of all the test components (cartridge, ID chip and detection buffer, diluent) must match each other.
- Do not interchange the test components between different lots

or use the test components after the expiration date, either of which might yield misleading of test result(s).

- Do not reuse. A detection buffer tube should be used for processing one sample only. So should a cartridge.
- The cartridge should remain sealed in its original pouch before use. Do not use the cartridge, if it is damaged or already opened.
- Frozen sample should be thawed only once. For shipping, samples must be packed in accordance with the regulations. Sample with severe hemolytic and hyperlipidemia cannot be used and should be recollected.
- Just before use, allow the cartridge, detection buffer and sample to be at room temperature for approximately 30 minutes.
- **ichroma™ Dengue NS1 Ag** as well as the instrument for ichroma™ tests should be used away from vibration and/or magnetic field. During normal usage, it can be noted that instrument for ichroma™ tests may produce minor vibration.
- Used detection buffer tubes, pipette tips and cartridges should be handled carefully and discarded by an appropriate method in accordance with relevant local regulations.
- An exposure to larger quantities of sodium azide may cause certain health issues like convulsions, low blood pressure and heart rate, loss of consciousness, lung injury and respiratory failure.
- **ichroma™ Dengue NS1 Ag** will provide accurate and reliable results subject to the following conditions.
 - Use **ichroma™ Dengue NS1 Ag** should be used only in conjunction with instrument for ichroma™ tests.
 - Any anticoagulants other than EDTA, heparin, sodium citrate should be avoided.

STORAGE AND STABILITY

- The cartridge is stable for 20 months (while sealed in an aluminum foil pouch) if stored at 4-30 °C.
- The detection buffer is stable for 20 months if stored at 2-8 °C.
- The diluent dispensed in a vial is stable for 20 months if stored at 2-8 °C.
- After the cartridge pouch is opened, the test should be performed immediately.

LIMITATION OF THE TEST SYSTEM

- The test may yield false positive result(s) due to the cross-reactions and/or non-specific adhesion of certain sample components to the capture/detector antigens.
- The test may yield false negative result. The non-responsiveness of the antibodies to the antigens is most common where the epitope is masked by some unknown components, so as not to be detected or captured by the antibodies. The instability or degradation of the antigen with time and/or temperature may cause the false negative as it makes antibody unrecognizable by the antigens.
- Other factors may interfere with the test and cause erroneous results, such as technical/procedural errors, degradation of the test components/reagents or presence of interfering substances in the test samples.
- Any clinical diagnosis based on the test result must be supported by a comprehensive judgment of the concerned physician including clinical symptoms and other relevant test results.

MATERIALS SUPPLIED

REF CFPC-62

Components of **ichroma™ Dengue NS1 Ag**

- Cartridge Box:

- Cartridges	25
- ID Chip	1
- Instruction For Use	1

- Aluminum Pouch containing Detection Buffer Tubes
 - Detection Buffer tubes 25
 - Dilution Buffer Vial Pouch
 - Diluent Vial 1

MATERIALS REQUIRED BUT SUPPLIED ON DEMAND

Following items can be purchased separately from **ichroma™ Dengue NS1 Ag**.

Please contact our sales division for more information.

- Instrument for ichroma™ tests
 - **ichroma™ II** REF FPRR021
- **ichroma™ Printer** REF FPRR007

SAMPLE COLLECTION AND PROCESSING

The sample type for **ichroma™ Dengue NS1 Ag** is human whole blood/serum/plasma.

- It is recommended to test the sample within 24 hours after collection.
- The serum or plasma should be separated from the clot by centrifugation within 3 hours after the collection of whole blood.
- Samples may be stored for up to a week at 2-8 °C prior to being tested. If testing will be delayed more than a week, samples should be frozen at -20 °C.
- Samples stored frozen at -20 °C for 2 months showed no performance difference.
- However, the whole blood sample should not be kept in a freezer in any case.
- Once the sample was frozen, it should be thawed one time and only for test, because repeated freezing and thawing can result in the changed test values.
- Samples containing precipitates must be clarified by centrifugation.

TEST SETUP

- Check the contents of **ichroma™ Dengue NS1 Ag**: Sealed Cartridge, Detection Buffer Tubes, Diluent Vial and ID Chip.
- Ensure that the lot number of the cartridge matches that of the ID chip, the detection buffer as well as the diluent.
- Keep the sealed cartridge (if stored in refrigerator), the detection buffer tube and the diluent at room temperature for at least 30 minutes just prior to the test. Place the cartridge on a clean, dust-free and flat surface.
- Turn on the Instrument for ichroma™ tests.
- Insert the ID Chip into the ID chip port of the Instrument for ichroma™ tests.
- Press the 'Select' button on the Instrument for ichroma™ tests. (Please refer to the 'Instrument for ichroma™ tests Operation Manual' for complete information and operating instructions.)

TEST PROCEDURE

- 1) Transfer 150 µL of diluent using a pipette to the detection buffer tube.
- 2) Transfer 75 µL of sample (whole blood/serum/plasma/control) using a pipette to the detection buffer tube.
- 3) Dissolve the granulated detection buffer thoroughly by tapping 10 times and pipetting 10-20 times.
(The sample mixture must be used immediately.)
- 4) Pipette out 75 µL of a sample mixture and dispense it into the sample well on the cartridge.
- 5) Leave the sample-loaded cartridge at room temperature for 12 minutes.
⚠ Scan the sample-loaded cartridge immediately when the incubation time is over. If not, it will cause incorrect test result.
- 6) To scan the sample-loaded cartridge, insert it into the cartridge holder of the instrument for ichroma™ tests. Ensure proper orientation of the cartridge before pushing it all the way inside

the cartridge holder. An arrow has been marked on the cartridge especially for this purpose.

- 7) Tab the 'Start' icon on the screen.
- 8) Instrument for ichroma™ tests will start scanning the sample-loaded cartridge immediately.
- 9) Read the test result on the display screen of the instrument for ichroma™ tests.

INTERPRETATION OF TEST RESULT

- Instrument for ichroma™ tests calculates the test result automatically and displays "Positive / Negative / Indeterminate".
- Ancillary value is served in the form of a cut-off index (COI).

Cut-off index (COI)	Result	Note
≤ 0.9	Negative for Dengue NS1 Ag	No need to additional test.
> 0.9, < 1.1	Indeterminate	Need to retest. If test results are shown 'Negative' or 'Indeterminate' repeatedly, these samples are considered dengue NS1 antigen negative.
≥ 1.1	Positive for Dengue NS1 Ag	Need to confirmation test.

QUALITY CONTROL

- Quality control tests are a part of the good testing practice to confirm the expected results and validity of the assay and should be performed at regular intervals.
- The control tests should be performed immediately after opening a new test lot to ensure the test performance is not altered.
- Quality control tests should also be performed whenever there is any question concerning the validity of the test results.
- Control materials are not provided with **ichroma™ Dengue NS1 Ag**. For more information regarding obtaining the control materials, contact **Boditech Med Inc.'s Sales Division for assistance**.
(Please refer to the instruction for use of control material.)

PERFORMANCE CHARACTERISTICS

- **Analytical Sensitivity**

- Limit of Detection (LOD)

To determine the limit of detection (LOD), the panel was tested by serial dilution with **ichroma™ Dengue NS1 Ag**.

Titre of cultured virus (pfu/mL)	Commercial Dengue qPCR kit		Commercial Rapid kit		ichroma™ Dengue NS1 Ag	
	Ct	Result	Band	Result	COI	Result
1.01 x 10 ⁶	19.20	Pos	+++	Pos	7.15	Pos
1.01 x 10 ⁵	22.92	Pos	++	Pos	8.6	Pos
1.01 x 10 ⁴	26.71	Pos	+	Pos	2.1	Pos
1.01 x 10 ³	30.50	Pos	-	Neg	1.3	Pos
1.01 x 10 ²	34.54	Pos	-	Neg	1.1	Pos
1.01 x 10 ¹	38.42	Pos	-	Neg	0.8	Neg
1.01 x 1	ND	Neg	-	Neg	0.98	Neg

- **Analytical Specificity**

- Cross-reactivity

There was no false positive result from 130 samples containing potentially interfering substances with the **ichroma™ Dengue NS1 Ag** test. The overall specificity was 100 %.

Clinical category	ichroma™ Dengue NS1 Ag		
	Number of samples	Negative	Positive
CHIKV	10	10	0
Zika (Uganda)	10	10	0
Zika (Panama)	10	10	0
Zika (Columbia)	10	10	0
JEV-1	10	10	0
JEV-3	10	10	0
HIV	10	10	0

HBV	20	20	0
CMV	20	20	0
EBV	20	20	0
Total	130	130	0

- Interference

There was no significant interference from these material with the **ichroma™ Dengue NS1 Ag test**.

Materials	Concentration
Heparin	100,000 U/L
K2_EDTA	5 µmol/L
Sodium citrate	25 mg/mL
Bilirubin	0.5 mmol/L
Hemoglobin	2 g/L
Triglycerides	1.5 mg/mL
Cholesterol	20 mmol/L
Albumin	30 mg/mL

■ **Precision**

- Between Lot

One person tested three different lots of **ichroma™ Dengue NS1 Ag**, ten times at each concentration of the control standard.

- Between person

Three different persons tested same lot of **ichroma™ Dengue NS1 Ag**, five times at each concentration of the control standard.

- Between day

One person tested same lot of **ichroma™ Dengue NS1 Ag** during three days, five times at each concentration of the control standard.

- Between site

One person tested same lot of **ichroma™ Dengue NS1 Ag** at three different sites, five times at each concentration of the control standard.

Dengue NS1 Ag Cal	Between lot		Between person		Between day		Between site	
	Positive/ Number of test	Positive	Positive/ Number of test	Positive	Positive/ Number of test	Positive	Positive/ Number of test	Positive
Negative	0/10	0 %	0/5	0 %	0/5	0 %	0/5	0 %
High	10/10	100 %	5/5	100 %	5/5	100 %	5/5	100 %
Mid	10/10	100 %	5/5	100 %	5/5	100 %	5/5	100 %
Low	10/10	100 %	5/5	100 %	5/5	100 %	5/5	100 %

■ **Comparability with reference product**

Commercial RT-PCR test				
ichroma™ Dengue NS1 Ag		Positive	Equivocal	Negative
	Positive	38	0	2
	Equivocal	0	0	0
	Negative	0	0	105
Total		40	0	107

- Sensitivity = $38/40 \times 100 = 95 \%$

- Specificity = $105/105 \times 100 = 100 \%$

REFERENCES

1. High circulating levels of the dengue virus nonstructural protein NS1 early in dengue illness correlate with the development of dengue hemorrhagic fever. Daniel H. Library et al., The Journal of Infectious Diseases, 2002.
2. Potential application of nonstructural protein NS1 serotype-specific immunoglobulin G enzyme-linked immunosorbent assay in the seroepidemiologic study of dengue virus infection: Correlation of results with those of the plaque reduction neutralization test. Pei-Yun S. et al., Journal of Clinical Microbiology, 2002.
3. Evaluation of diagnostic tests: Dengue, Rosanna W. P. et. al., Nature, 2010.

Note: Please refer to the table below to identify various symbols.

	Sufficient for <n> tests
	Read instruction for use
	Use by Date
	Batch code
	Catalog number
	Caution
	Manufacturer
	Authorized representative of the European Community
	In vitro diagnostic medical device
	Temperature limit
	Do not reuse
	This product fulfills the requirements of the Directive 98/79/EC on in vitro diagnostic medical devices

CE