



ichroma™ IL-6

INTENDED USE

ichroma™ IL-6 is a fluorescence immunoassay (FIA) for the quantitative detection of IL-6 in human whole blood /serum/plasma. It is helpful as an aid in management and monitoring of inflammatory disease.

For *in vitro* diagnostic use only.

INTRODUCTION

IL-6 (Interleukin-6) is produced by a variety of cells including T cells, B cells, fibroblasts, endothelial cells, monocytes, keratinocytes, mesangial cells, and some tumor cells. The genes for human and murine IL-6 have been cloned and sequenced. Human IL-6 has a molecular mass of 21 to 28 kDa and is comprised of 212 amino acids that include two possible N-glycosylation sites and four cysteine residues.

IL-6 is a pleiotropic cytokine with multiple roles in the regulation of inflammation and hematopoiesis. IL-6 is produced at the site of inflammation and plays a key role in the acute phase response as defined by a variety of clinical and biological features such as the production of acute phase proteins.

IL-6 is the major regulator of the acute phase response in human hepatocytes. Due to its pleiotropic action, IL-6 has been intensively studied in many laboratories. It turned out to be an important factor in the immune and in the hematopoietic system and the major mediator in the hepatic acute phase response.

IL-6 is one of the proinflammatory cytokines and is detected in serum in the early stages of infections. Particularly in bacterial infections, IL-6 levels may be higher than CRP in early disease stages, and this may be helpful for early diagnosis. Early in infection, the CRP level may be low, but serial measurements can provide useful results and can be helpful in deciding when to discontinue antibiotic treatment. The combination of IL-6 and CRP has recently been proven to be useful in the early diagnosis of sepsis in newborns.

PRINCIPLE

This test uses a sandwich immunodetection method.

The detector antibodies in buffer bind to antigens in the sample, forming antigen-antibody complexes, and migrate onto nitrocellulose matrix to be captured by the other immobilized- antibodies on a test strip.

More antigens in the sample will form more antigen-antibody complexes which lead to stronger fluorescence signal by detector antibodies, which is processed by the instrument for ichroma tests to show IL-6 concentration in the sample.

COMPONENTS

ichroma™ IL-6 consists of 'cartridges', 'detector tubes',

'detector diluent'.

- The cartridge contains the membrane called a test strip which has streptavidin at the test line, and chicken IgY at the control line. All cartridges are individually sealed in an aluminum foil pouch containing a desiccant, and they are further packaged in a box.
- The detector tube has 2 granules containing anti-chicken IgY fluorescence conjugate, anti-Interleukin 6 fluorescence conjugate, anti-Interleukin 6 biotin conjugate, HAMA Blocker, bovine serum albumin(BSA), bromophenol blue and sucrose as a stabilizer and sodium azide as a preservative in Tris buffer and phosphate buffered saline. All detector tubes are packed in a pouch.
- The detector diluent contains sodium chloride, Tween20, TTAB, sodium azide as a preservative in Tris buffer and it is pre-dispensed in vial. The detector diluent is packed in a box.

WARNINGS AND PRECAUTIONS

- For *in vitro* diagnostic use only.
- Follow the instructions and procedures described in this 'Instructions for use'.
- Use only fresh samples and avoid direct sunlight.
- Lot numbers of all the test components (cartridge, detector tube, detector diluent, capillary tube, and ID chip) must match each other.
- Do not interchange the test components between different lots or use the test components after the expiration date, either of which might yield incorrect test result(s).
- Do not reuse cartridge or detector tubes. A cartridge should be used for testing one sample only. A detector tube should be used for processing of one sample only.
- The cartridge should remain sealed in its original pouch until just before use. Do not use the cartridge, if pouch is damaged or has already opened.
- Frozen sample should be thawed only once. For shipping, samples must be packed in accordance with local regulations. Sample with severe hemolysis and/or hyperlipidemia must not be used.
- If test components and/or sample are stored in refrigerator, then allow cartridge, detector tube, detector diluent and sample to be at room temperature for approximately 30 minutes before use.
- The instrument for ichroma™ tests may generate slight vibration during use.
- Used cartridges, detector tubes, detector diluent, capillary tubes and pipette tips should be handled carefully and discarded by an appropriate measure in accordance with relevant local regulations.
- The detector tube and the detector diluent contain sodium azide (NaN₃), and they may cause certain health issues like convulsions, low blood pressure and heart rate, loss of consciousness, lung injury and respiratory failure. Avoid contact with skin, eyes, and clothing. In case of contact, rinse immediately with running water.
- No Biotin interference was observed in ichroma™ IL-6 when biotin concentration in the sample was below 5 ng/mL. If a patient has been taking biotin at dosage of more than 0.03 mg a day, it is recommended to test again 24 hours after discontinuation of biotin intake.
- **ichroma™ IL-6** will provide accurate and reliable results

subject to the below conditions.

- **ichroma™ IL-6** should be used only in conjunction with instrument for ichroma™ tests.
- Have to use recommended anticoagulant.

Recommended anticoagulant

K₂ EDTA, K₃ EDTA, Lithium heparin

- The capillary tube should be used when the following conditions are met.
 - The capillary tube provided with the kit is recommended to obtain correct test result.
 - Whole blood should be immediately tested after collection.
 - Excess whole blood around the capillary tube should be wiped off.
 - In order to avoid cross-contamination, please do not re-use capillary tube for multiple samples.

LIMITATION OF THE TEST SYSTEM

- The test may yield false positive result(s) due to the cross-reactions and/or non-specific adhesion of certain sample components to the capture/detector antibodies.
- The test may yield false negative result(s) due to the non-responsiveness of the antigen to the antibodies which is the most common if the epitope is masked by some unknown components, so therefore not being able to be detected or captured by the antibodies. The instability or degradation of the antigen with time and/or temperature may also cause false negative result as it makes antigen unrecognizable by the antibodies.
- Other factors may interfere with the test and cause erroneous results, such as technical/procedural errors, degradation of the test components/reagents or presence of interfering substances in the test samples.
- Any clinical diagnosis based on the test result must be supported by a comprehensive judgment of the concerned physician in conjunction with clinical symptoms and other relevant test results.

STORAGE AND STABILITY

Storage condition			
Component	Storage Temperature	Shelf life	Note
Cartridge	2 - 30 °C	20 months	Disposable
Detector tube	2 - 30 °C	20 months	Disposable
Detector diluent	2 - 30 °C	20 months	Unopened
	2 - 30 °C	20 months	Opened

- After the cartridge pouch is opened, the test should be performed immediately.

MATERIALS SUPPLIED

REF CFPC-116

Components of **ichroma™ IL-6**

- Cartridge Box
 - Cartridge 25
 - Detector tube 25
 - Detector diluent 1
 - 35 µL Capillary tube 25
 - ID chip 1
 - Instructions for use 1

MATERIALS REQUIRED BUT SUPPLIED ON DEMAND

Following items can be purchased separately from **ichroma™ IL-6**.

Please contact our sales division for more information.

- **ichroma™ II** **REF** FPRR021
- **ichroma™ III** **REF** FPRR037
- **ichroma™ M2** **REF** FPRR031
- **ichroma™ 50** **REF** FPRR022
- **ichroma™ -50 PLUS** **REF** FPRR036
- **Boditech IL-6 Control** **REF** CFPO-296

SAMPLE COLLECTION AND PROCESSING

The sample type for **ichroma™ IL-6** is human whole blood /serum/plasma.

- It is recommended to test the sample within 24 hours after collection.
- The samples (serum, plasma) should be separated from the clot by centrifugation within 3 hours after the collection of whole blood.
- The samples (whole blood, serum, plasma) may be stored for a week at 2-8 °C prior to being tested. If testing will be delayed more than a week, samples(except the whole blood) should be frozen at below -20 °C.
- The samples (serum, plasma) stored frozen at -20 °C for 3 months showed no performance difference.
- However, the whole blood sample should not be kept in a freezer in any case.
- As a repeated freeze-thaw cycle may affect the test result, do not refreeze previously frozen samples.

[35 µL Capillary tube]

Whole blood sample may be used to collect according to below:

- ① Wear disposable gloves and protective equipment for safety.
- ② Open the zipper bag which has capillary tubes.
- ③ Take out the capillary tube and check for damage or contamination.
- ④ Hold the handle of the capillary tube and touch the surface of blood with the capillary tube.
- ⑤ Fill it with blood completely.
(Make sure that no air bubbles are present in the capillary tube. Do not get blood on the surface of the capillary tube. If the blood gets on the surface of the capillary tube, remove it gently with gauze.)

TEST SETUP

- Check the contents of **ichroma™ IL-6**: Sealed cartridges, detector tubes, detector diluent, capillary tubes, an ID chip and an Instructions for use.
- Ensure that the lot number of the cartridge matches that of the detector tube, detector diluent as well as an ID chip.
- If the sealed cartridge, the detector tube and the detector diluent have been stored in a refrigerator, place them on a clean and flat surface at room temperature for at least 30 minutes before testing.
- Turn on the instrument for ichroma™ tests.
- Insert the ID chip into the 'ID chip port'.
(Please refer to the instrument for ichroma™ tests operation manual for complete information and operating instructions)

TEST PROCEDURE

▶ **ichroma™ II, ichroma™ M2**

Multi test mode / Read now mode

- 1) Take 150 µL of the detector diluent using a pipette and dispense it to the detector tube containing a granules. When the granule form is completely dissolved in the tube, it becomes detection buffer. (The detection buffer must be used immediately. Do not exceed 30 seconds.)
- 2) Take 35 µL of sample (human whole blood/serum/plasma/control) using a pipette and dispense it to the detector tube.
 - ※ If you use a capillary tube (35 µL), put it into the detector tube after collecting sample.
- 3) Close the lid of the detector tube and mix the sample thoroughly by shaking it about 20 times. (The sample mixture must be used immediately. Do not exceed 30 seconds.)
- 4) Take 75 µL of the sample mixture and dispense it into the sample well of the cartridge.
- 5) Leave the cartridge at room temperature for 12 minutes.
 - ⚠ Scan the sample-loaded cartridge immediately when the incubation time is over. If not, it will cause inaccurate test result.
- 6) To scan the sample-loaded cartridge, insert it into the cartridge holder of the instrument for ichroma™ tests. Ensure proper orientation of the cartridge before pushing it all the way inside the cartridge holder. An arrow is marked on the cartridge especially for this purpose.
- 7) Press the 'Select' or tap the 'Start' button on the instrument for ichroma™ tests to start the scanning process. (ichroma™ M2 is tested automatically after inserting.)
- 8) The instrument for ichroma™ tests will start scanning the sample-loaded cartridge immediately.
- 9) Read the test result on the display screen of the instrument for ichroma™ tests.

Single test mode/ Walk away mode

- 1) The test procedure is same with the 'Multi test mode 1) - 4)'.
- 2) Insert the sample-loaded cartridge into the holder of the instrument for ichroma™ tests. Ensure proper orientation of the cartridge before pushing it all the way inside the cartridge holder. An arrow is marked on the cartridge especially for this purpose.
- 3) Press the 'Select' or tap the 'Start' button on the instrument for ichroma™ tests. (ichroma™ M2 is tested automatically after inserting.)
- 4) The cartridge goes inside the instrument for ichroma™ tests and will automatically start scanning the sample-loaded cartridge after 12 minutes.
- 5) Read the test result on the display screen of the instrument for ichroma™ tests.

▶ **ichroma™ III**

- 1) The test procedure is same with the 'Single test mode'.

▶ **ichroma™-50, ichroma™-50 Plus**

- 1) Insert the tip array in the tip station.
- 2) Insert the detector tube in the reagent station and cover the reagent station to hold the detector tubes in place.
- 3) Open the lid of the detector diluent and insert the detector diluent in the diluent station.
- 4) Insert the cartridge magazine with the cartridges into the magazine station.
- 5) Insert the sample tube into the blood collection tube rack and load the blood collection tube rack into the sampling station (loading part).
- 6) Tap the button located in the upper side of the No. of test cartridge region to select the ID chip that you want to use.
- 7) When the selected cartridge slot is activated, set the number of the detector tube by tapping.
- 8) Set the number of pipette tips by tapping.
- 9) Tap the 'Start' button on the left upper of the main screen to start test.

INTERPRETATION OF TEST RESULT

- The instrument for ichroma™ tests calculates the test result automatically and displays IL-6 concentration of the test sample in terms of pg/mL.
- Reference value : 7 pg/mL
- Working range : 2 - 2,500 pg/mL

QUALITY CONTROL

- Quality control tests are a part of the good testing practice to confirm the expected results and validity of the assay and should be performed at regular intervals.
- Quality control tests should also be performed whenever there is any question concerning the validity of the test results.
- Control materials are provided on demand with **ichroma™ IL-6**. For more information regarding obtaining the control materials, contact Boditech Med Inc.'s Sales Division for assistance. (Please refer to the instructions for use of control material.)

PERFORMANCE CHARACTERISTICS

■ **Analytical sensitivity**

LOB (Limit of Blank)	0.5 pg/mL
LOD (Limit of Detection)	1.0 pg/mL
LOQ (Limit of Quantitation)	2.0 pg/mL

■ **Analytical specificity**

[Cross-reactivity]

Biomolecules listed in the following table were added to the test sample(s) at concentrations much higher than their normal physiological levels in the blood. **ichroma™ IL-6**—test results did not show any significant cross-reactivity with these biomolecules.

Cross reactivity materials	Conc.
Interleukin-1α	50 ng/mL
Interleukin-1β	50 ng/mL
Interleukin-2	50 ng/mL
Interleukin-3	50 ng/mL
Interleukin-4	50 ng/mL
Interleukin-8	50 ng/mL

Interferon-γ	50 ng/mL
TNF-α	50 ng/mL

[Interference]

Interferents listed in the following table were added to the test sample at the concentration mentioned below. **ichroma™ IL-6** test results did not show any significant interference with these materials.

Interference materials	Conc.
Bilirubin	342 μmol/L
Cholesterol	13 mmol/L
D-Glucose	55 mmol/L
Hemoglobin	2 g/L
L-Ascorbic acid	170 μmol/L
Triglyceride	37 mmol/L
EDTA	3.4 μmol/L
Heparin	3,000 U/L

■ Precision

[Single-site study]

- Repeatability (within-run precision)
- within-laboratory precision (Total precision)
- Lot to lot precision

3 Lots of **ichroma™ IL-6** were tested for 20 days. Each standard material was tested 2 times per day. For each test, each material was duplicated.

Conc. [pg/mL]	Repeatability (within-run)		within-laboratory		Lot to lot precision	
	Mean [pg/mL]	CV (%)	Mean [pg/mL]	CV (%)	Mean [pg/mL]	CV (%)
9.00	9.10	5.7	9.11	5.8	9.07	6.1
42.61	43.12	6.2	42.81	6.4	43.07	6.2
1,274	1,271.98	6.3	1,272.96	6.3	1,272.35	6.4

[Multi-site study]

- Reproducibility

1 Lot of **ichroma™ IL-6** was tested for 5 days in 3 different sites (1 person per 1 site, 1 instrument per 1 site).

Each standard material was tested 1 time per and 5 replicates per day.

Multi-site study			
Conc. [pg/mL]	Reproducibility		
	Mean [pg/mL]	SD	CV (%)
9.00	9.22	0.55	5.9
42.61	42.28	2.89	6.8
1,274	1,272.83	73.61	5.8

■ Accuracy


The accuracy was confirmed by testing with 3 different lots of **ichroma™ IL-6**. The tests were repeated 10 times each concentration of the control standard.

Expected value [pg/mL]	Mean [pg/mL]			3 LOTS of AVG [pg/mL]	Recovery (%)
	Lot 1	Lot 2	Lot 3		
9.00	9.02	8.89	8.69	8.87	98.5
254.80	249.81	257.58	252.00	253.13	99.3
509.60	494.62	472.06	502.36	489.68	96.1
764.40	730.64	739.38	746.37	738.80	96.7
1,019.20	1,005.76	967.84	1,000.54	991.38	97.3
1,274.00	1,252.09	1,256.52	1,292.03	1,266.88	99.4

■ Comparability

IL-6 concentrations of 110 clinical samples were quantified independently with **ichroma™ IL-6** (**ichroma™ II**) and Comparator A as per prescribed test procedures. Test results were compared and their comparability was investigated with linear regression and correlation coefficient (R). The regression equation and correlation













coefficient are as follows.

**REFERENCES**

1. Interleukin 6: From bench to bedside. Nishimoto N, Kishimoto T. Nat Clin Pract Rheumatol. 2006 Nov;2(11):619-26.
2. The biology of interleukin-6. Kishimoto T. Blood. 1989 Jul;74(1):1-10.
3. Interleukin-6: An overview. Van Snick J. Annu Rev Immunol. 1990;8:253-78.
4. Complementary DNA for a novel human interleukin (BSF-2) that induces B lymphocytes to produce immunoglobulin. Hirano T, Yasukawa K, Harada H, et al. Nature. 1986 Nov;324(6092):73-6.
5. Interleukin-6 and chronic inflammation. Cem Gabay. Arthritis Research & Therapy 2006, 8(Suppl 2):S3
6. Interleukin-6 and the acute phase response. Peter C. HEINRICH, Jose V. CASTELL and Tilo ANDUS. Biochem. J. (1990) 265, 621-636
7. Interleukin-6: A sensitive parameter for the early diagnosis of neonatal bacterial infection. Buck C, Bundschu J, Gallati H, Bartmann P, Pohlandt F. Pediatrics 1994;93:54-58.
8. Evaluation of IL-6, TNF-alpha and IL-1 beta for early diagnosis of neonatal sepsis. Silveria RC, Procionay RS. Acta Paediatr 1999;88:647-650.
9. Significance of serial C-reactive protein responses in neonatal infection and other disorders. Pourcyrous M, Bada HS, Korones SB, Baselski V, Wong SP. Pediatrics 1993;92:431-435.
10. Interleukin-6 concentrations in neonates evaluated for sepsis. Doellner H, Arntzen KJ, Haereid PE, Aag S, Austgulen R. J Pediatr 1998;132:295-299.
11. What are the Cut-Off Levels for IL-6 and CRP in Neonatal Sepsis? Istemi Han Celik, Fatma Gamze Demirel, Nurdan Uras, Serife Suna Oguz, Omer Erdevce, Zeynep Biyikli, and Ugur Dilmen. Journal of Clinical Laboratory Analysis 24 : 407-412 (2010)

12.

Note: Please refer to the table below to identify various symbols.

	Sufficient for <n> tests
	Read instruction for use
	Use by Date
	Batch code
	Catalog number
	Caution
	Manufacturer
	Authorized representative of the European Community
	In vitro diagnostic medical device
	Temperature limit
	Do not reuse
	This product fulfills the requirements of the Directive 98/79/EC on in vitro diagnostic medical devices